JS-VSG-10 Voltage Signal Generator



User Manual

V202308

1 Features:

- 1.1 Adjustable output of 0-12.5V, and the output range can be set arbitrarily within 0-12.5 V;
- 1.2 -1999 to 9999, decimal point position can be set arbitrarily;
- 1.3 4-bit LCD (with backlight), digital encoder knob tuning (the number of turns can be set);
- 1.4 Manual tuning or programmable automatic output, can dynamically output continuous curve (set with encoder knob with key combination);
- 1.5 Can set fast switching coarse tuning and fine tuning mode, fixed startup value mode, fast return to zero and other modes;
- 1.6 The output can be calibrated and the error can be corrected linearly;
- 1.7 Output short circuit protection, power supply reverse connection protection, etc.;
- 1.8 Industrial grade circuit design, can work without power for a long time;

2 Technical Indicators:

- 2.1 Power supply DC 8-28V/1W (note that power supply must be 2V larger than output, such as 0-10V, power supply must be greater than 12V);
- 2.2 Output range: 0-12.5V Maximum current: 20mA; Tuning display accuracy: 0.01 V error < 0.03 V
- 2.3 Encoder knob pulse number of 20, Segment LCD;
- 2.4 Working environment:-20-60°C, relative humidity < 80%;

3 Dimension Drawing:



Attention for the installation of cabinet/electric box: The panel must be stuck to the ears on both sides to fix it, so the thickness of the panel must be greater than 1.4 mm, The opening size should consider the width of the ear, and should not be too small, otherwise it will not fit in. The recommended opening size is 77X40mm

4 Wiring Diagram:



5 Parameter Settings:

(Press the knob to confirm ("OK"), rotated clockwise is "+", and counterclockwise is "-"):

- 5.1 Press the knob for 2 seconds to enter the parameter setting state, "F001", then press the knob to set the value, and press again to save after modification;
- 5.2 For the parameters after F002, you need to enter the password. After entering the setting display F001, display 4 horizontal bars clockwise, and then enter:
- 5.3 To enter F002 ... enter the password "+ - +" first;
- 5.4 To enter F200 ... enter the password "- + +" (automatic curve output setting);
- 5.5 Rotate the knob directly to the last parameter number, press the "OK", and enter the normal operation screen after setting;
- 5.6 After the parameter setting screen has no operation for more than 10s, it exits the setting state and enter the normal operation screen;
- 5.7 F006/7/9/10、Ft01-9、FA01-9、Fb01-9 after entering the value setting of these parameters, one of the nixie tube will blink, press and hold the knob for 2 seconds to change the blinking position of the nixie tube, and you can switch the adjustment bit;

6 Parameter Table and Description:

No.	Description	Remarks	Defau					
			lt					
F001	Coarse or fine tuning	0: Coarse tuning mode, "F002" to modify the addition and	0					
		ubtraction multiples						
		1: Fine tuning mode, "F003" to modify addition and						
		subtraction multiples						

-			
		2: Automatic curve output (parameter $F201 > 0$ should be set	
		first) (for burn-in test products)	
F002	Coarse tuning of	1-100 (x 10)	1
	addition and subtraction		
	multiples		
F003	Fine tuning of	1-100	1
	addition and subtraction		
E004	multiples		0
F004	Press function	0: Manually store the output value (fixed startup value);	0
		1. Quickly switch coarse tuning and fine tuning;	
		2: Output OFF/ON;	
		3. Quick return to zero (minimum value);	
		(Function 1-3 is automatic storage of output value: 3 seconds	
FOOF	Output range (V)	$0.0 \ 10V \ 1.2 \ 10V \ 2.0 \ EV \ 2.1 \ EV \ 4.0 \ 2 \ 2V \ E.0 \ 2 \ EV \ 6.0 \ 1V$	
F005	Output range(v)	7.0\/ 12\/ 1: Cuctom	
E006	Custom output low and	0.12.50V	0
F007	Custom output low end	0.12.50V	10.00
F008	Display mode	0: Actual voltage 1: 0.100 0(%) 2:0.50 0(bz) 3:0.90 4:0.100	10.00
1008	Display mode	5.0-200	
		6·0-400 7·0-600 8·0-1000 9·0-1300 10·0-1500 11·0-2000	
		12.0-2500 -1: Custom	
F009	Custom display low end	-1999 to 9999, the decimal point set in F011	0
F010	Custom display high	-1999 to 9999, the decimal point set in F011	1000
	end	For the second product of the second s	
F011	Custom decimal point	0-4 0/1: None 2: 999.9 3: 99.99 4: 9.999	3
	position		
F012-	Standby	Standby	
14			
F015	+ 10V calibration value	-999+999 for internal reference only, please be careful when	
		modifying	
F200	Automatic curve mode	0: automatic cycle	0
		1: curve run once when power on, then output OV, and then	
		press the knob to run again	
		2:curve run once when power on, then output last curve value,	
		and then press the knob to run again	
		3:=mode 1, but not run when power on	
		4: =mode 2, but not run when power on	
F201	Curve number	0: Automatic curve output mode does not need; 1-9: number of	0
		sections	
Ft01	Section 1 curve time	0-999 seconds Set as many values as there are sections of	
		"F201"	
FA01	Section 1 starting	0.00-12.50 V	
	voltage		

Fb01	Section 1 end voltage	0.00-12.50 V	
Ft02	Section 2 curve time	0-999 seconds	
Fb09	Section 9 end voltage	0.00-12.50 V	

6.1 Examples of setting and calculation of knob turns:

Press the knob for 2 seconds, enter the setting, display F001, and then press it to set its value to 0 (coarse tuning) or 1 (fine tuning), which can quickly switch the tuning speed, and the multiples of coarse tuning and fine tuning are set in F002 and F003;

Examples of number of turns calculated: Knob encoder 20 grids per turn

Setting example	F001	F002	F003	Description
0-10V shows 0-10.00, and the knob is	0	5	Х	Set coarse tuning 5, with a grid
adjusted for 1 turn				change of 0.5V
0-10V shows 0-10.00, and the knob is	1	X	1	Set fine tuning 1, with a grid
adjusted for 50 turns				change of 0.01V
0-5V shows 0-5.00, and the knob is adjusted	1	X	5	Set fine tuning 5, with a grid
for 5 turns				change of 0.05V

Save the power-on value when it is powered down. Press the knob to set other functions: 6.2

F004=0: After adjusting the knob, press the knob to save it, and save as much as you turn it on;

F004=1: Press the knob to switch the manual tuning speed, which is equal to setting F001=0 or 1;

F004=2: Short press knob, switch output, OFF state output is 0V;

F004=3: Press the knob, and the screen display value is directly adjusted to minimum display value;

		T 00 T	Energia		T 000	
63	Examples of output	it range a	and display	v scale sett	ings.	

Setting example	F005	F006	F007	F008	F009	F010	F011
0-10V show 0-10.00	0	x	x	0	x	Х	x
0-10V show 0-100.0	0	х	X	1	х	Х	Х
0-10V show 0-50.0	0	х	x	2	х	х	х
0-10V show 0-5000	0	х	х	-1	0	5000	0
0-5V show 0-5.00	2	x	x	0	х	х	x
0-5V show 0-100.0	2	х	x	1	х	х	х
0-5V show 0-50.0	2	x	x	2	х	х	х
0-3.3V show 0-3.30	4	х	х	0	х	х	х
0-3.3V show 0-100.0	4	х	х	1	х	х	х
0-3.3V show 0-50.0	4	х	х	2	х	х	х
5-10V show 5.00-10.00	-1	5.00	10.00	-1	500	1000	3
3-5V show -80 to 1000	-1	3.00	5.00	-1	-80	1000	0
1-3V show 1.00-3.00	-1	1.00	3.00	-1	100	300	3
2-8V show 0-250	-1	2.00	8.00	-1	0	250	0
0-12V show 0-12.00	7	x	x	0	x	х	x
0-12.5V show 0-12.50	-1	0	12.50	0	х	Х	х
0-2V show 0-2.000 (1mV)	-1	0.0	2.00	-1	0	2000	4
0-10V show 0-10.000							
(1mV Adjust)	2	х	х	-1	0	A000	4

Note 1: "10.000" cannot be displayed with 4 nibbles, so "A000" is used instead

6.4 Output error calibration method:

When there is an error between the meter display value and the multimeter measurement value, you can calibrate 10V and make the meter display consistent with the multimeter through linear correction;

Enter the parameter F015 setting, adjust its value, so that the multimeter measurement shows 10.00 V, press the knob to save, and the calibration is completed (the calibration value is an internal correction value regardless of the size);

6.5 Examples of automatic curve loop output setting steps: (For aging test products, enter the F200 password "- + - +")

Step 1. Set F200 curve run mode;

Step 2. Set F201 = number of sections, with a maximum of 9 sections, and automatically change the cycle output for aging TEST products..;

Step 3, setting each section: FtXX= time 1-999 sec/FAXX= start voltage/FbXX= end voltage;

Step 4, finally setting F001=2, switching from manual tuning mode to automatic curve output mode;

Turn off automatic loop output mode: F001=0 or 1;

Example of	Mode	Number of	Section 1	Section	Section	Section		
waveform		sections		2	3	4		
Triangular	F200 = 0	F201 = 2	Ft01 = 10	Ft02=10				
wave			FA01 = 2.00	FA02=9.				
			Fb01=9.00	00				
				Fb02=2.				
				00				
Square wave	F200 = 0	F201 = 2	Ft01 = 10	Ft02=10				
			FA01 = 6.00	FA02=3.				
2 0 t 1 →			Fb01=6.00	00				
				Fb02=3.				
				00				
Sine wave	F200 = 0	F201 = 6	Ft01 = 4	Ft02=3	Ft03=3	Ft01 =	Ft01=3	Ft01=3
10			FA01 = 3.00	FA02=5.	FA03=6.	4	FA01 =	FA01 =
			Fb01=5.00	00	00	FA01 =	3.00	2.00
				Fb02=6.	Fb03=5.	5.00	Fb01=2.	Fb01=3.
				00	00	Fb01=3.	00	00
						00		

7 Attention

- 7.1 Turn off the power supply before wiring;
- 7.2 Exceeding the range shown in the technical index, it may cause the instrument to work abnormally or even be damaged;